

ASAL
ORIGINAL

PIHAK BERKUASA
PERANTI PERUBATAN



MEDICAL DEVICE
AUTHORITY

PIHAK BERKUASA PERANTI PERUBATAN
MEDICAL DEVICE AUTHORITY
AKTA PERANTI PERUBATAN 2012 (AKTA 737)
MEDICAL DEVICE ACT 2012 (ACT 737)
SIJIL PENDAFTARAN PERANTI PERUBATAN
MEDICAL DEVICE REGISTRATION CERTIFICATE
Seksyen 5(1) Akta 737
Section 5(1) of Act 737

No. Pendaftaran: **GA10738125-203080**
Registration No.:

Tarikh Sah Pendaftaran:
Registration Validity Date:

15/05/2025 - 14/05/2030

Sijil ini adalah dengan ini diberi kepada: **RAINBOW OCULAR DEVICES SDN BHD**
This certificate is hereby issued to:

yang beralamat di:
which is located at:

**78A, JALAN BP7/8 BANDAR BUKIT PUCHONG,
47120 PUCHONG, SELANGOR,
47120 PUCHONG
SELANGOR DARUL EHSAN**

Peranan establismen
Role of establishment

Wakil Diberi Kuasa
Authorized Representative

bagi mengesahkan peranti perubatan seperti yang dinyatakan dalam Lampiran 1 adalah berdaftar di bawah Seksyen 5(1) Akta 737.

to confirm that the medical device as detailed out in Attachment 1 is registered under Section 5(1) of Act 737.

Pendaftaran ini diberikan tertakluk kepada peruntukan-peruntukan di bawah Akta 737 dan peraturan-peraturan yang dibuat dibawahnya serta syarat-syarat seperti di Lampiran 2.

This registration is granted subject to the provisions under Act 737 and its subsidiary legislations and the conditions as in Attachment 2.



MURALITHARAN PARAMASUA
KETUA EKSEKUTIF
CHIEF EXECUTIVE
PIHAK BERKUASA PERANTI PERUBATAN
MEDICAL DEVICE AUTHORITY



No. Pendaftaran: **GA10738125-203080**
Registration No.:

Tarikh Sah Pendaftaran: **15/05/2025 - 14/05/2030**
Registration Validity Date:

Sijil ini adalah dengan ini diberi kepada: **RAINBOW OCULAR DEVICES SDN BHD**
This certificate is hereby issued to:

Butir-butir peranti perubatan yang didaftarkan
Particulars of the registered medical device

Nama Peranti Perubatan **Fluorescein Sodium Ophthalmic Strips USP**
Medical Device Name

Kelas **CLASS A** Jenama **Fluoro Touch**
Class *Brand*

Kelompok **FAMILY**
Group

Nama dan alamat pembuat: **MADHU INSTRUMENTS PVT, LTD..
F-90/3D, OKHLA INDUSTRIAL AREA PHAS-I, NEW DELHI 110020 INDIA.**
Name and address of manufacturer **110020, INDIA**

APPENDIX

NO	NAME AS PER DEVICE LABEL	IDENTIFIER	BRIEF DESCRIPTION OF ITEM
1	Fluoro Touch 100's	MIPL/A4	100s in packaging
2	Fluoro Touch 300's	MIPL/A4	300s in packaging
"End Of Product List"			

DIAGNOSTIC STRIPS

Fluoro Touch

Fluorescein Sodium
Ophthalmic Strips

Most commonly used diagnostic strip to visualize defects or abrasions in the corneal epithelium by staining the areas of cellular loss.



Product Code	Description	Qty. per box
MIPL/A4	Fluorescein Sodium Ophthalmic Strips (Box of 100)	100 strips (Sterile)
	Fluorescein Sodium Ophthalmic Strips (Box of 300)	300 strips (Sterile)

Indications

- Useful for evaluating hard contact lens fitting
- Applanation tonometry
- Evaluating tear film breakup time
- To visualise defects in the corneal epithelium

Specifications

- Available as sterile strips in individual pouches
- Each strip is impregnated with 1.0 mg of Fluorescein sodium
- Innovative peel open packing pouch with transparent film on one side
- CE certified variant also available for evaluating contact lens fitting